



Revista Portuguesa de
Cardiologia
 Portuguese Journal of **Cardiology**
www.revportcardiol.org



EDITORIAL COMMENT

Catheter ablation of atrioventricular nodal re-entrant tachycardia: Humans versus machines?



Ablação por cateter de taquicardia por re-entrada auriculoventricular nodal: seres humanos *versus* máquinas?

António Creta^{a,b}, Rui Providência^{a,c,*}

^a Barts Heart Centre, St Bartholomew’s Hospital, Barts Health NHS Trust, London, United Kingdom
^b Campus Bio-Medico, Unit of Cardiology, University of Rome, Rome, Italy
^c Institute of Health Informatics Research, University College of London, London, United Kingdom

Following advances in ablation techniques and in cardiac electrophysiology in the late 1980s and early 1990s,^{1–3} catheter ablation of one of the pathways involved in the re-entrant mechanism of atrioventricular nodal re-entrant tachycardia (AVNRT) has now become part of modern cardiology. The 2015 American College of Cardiology/American Heart Association/Heart Rhythm Society guideline states that “catheter ablation of the slow pathway is recommended” for the ongoing management of AVNRT (class I recommendation, level of evidence B),⁴ and the latest European Society of Cardiology recommendations (a joint European and American guideline from 2003) are also clear on the role of ablation for recurrent symptomatic AVNRT, giving it a class I recommendation, level of evidence B.⁵

Slow pathway ablation was adopted worldwide during the 1990s, and due to the high success rate of this procedure (>99% reported by some highly experienced centers), the cardiac electrophysiology community assumed that it was ‘case closed’ for AVNRT. Subsequently, after the landmark

publication by Haissaguerre et al. in the late 1990s,⁶ attention shifted to atrial fibrillation, and AVNRT began to be considered a ‘simple arrhythmia’, and catheter ablation of the slow pathway a simple procedure.

However, this is by no means a risk-free procedure, and it should made clear to electrophysiology trainees beginning to acquire experience in catheter manipulation that not only must they become expert in interpreting signals and performing the diagnostic maneuvers to confirm a diagnosis of AVNRT, but they must also develop the skill to keep the catheter in a stable position, preventing it from migrating during the ablation part of the procedure, and maintaining sufficient contact force throughout the application to deliver an effective lesion. Also, they need a very quick ‘endocavitary-signals-to-brain connection’ to immediately abort application in the event of fast junctional rhythm or at the earliest signs of atrioventricular conduction damage. In this procedure, a minor mistake can lead to lifelong consequences (need for permanent pacemaker) in patients who are frequently young. In the hands of highly experienced centers and operators, the incidence of complete atrioventricular block has been reported to be 0.4%,⁷ but in the real world the incidence of this complication may in fact be much higher.

DOI of original article: <https://doi.org/10.1016/j.repc.2018.07.006>

* Corresponding author.
 E-mail address: rui_providencia@yahoo.com (R. Providência).

Acknowledging these aspects of the procedure is proof that research in the area is not over. Debate is ongoing regarding the most effective and safest ablation energy (cryoablation or radiofrequency),⁷ whether three-dimensional mapping systems should be routinely used⁸ (and when they are, whether there is a role for substrate mapping⁹⁻¹¹), zero-fluoroscopy procedures versus very short screening times,¹² and the potential role of contact force sensing for slow pathway ablation.¹³

However, a different question should be asked: are human beings good enough to manipulate the catheters? Is manual catheter manipulation preferable, or are alternatives like remote magnetic navigation systems (MNS) a better and safer approach?

Further evidence on this subject of the battle between machines and humans is elegantly provided by Parreira and colleagues in this issue of the *Journal*.¹⁴ Unlike previous studies in which comparisons involved very small samples, short follow-up durations and even catheters which are no longer in use, this study performs a careful comparison of two different ablation approaches using current technology in the hands of an experienced operator. Remote magnetic navigation ablation with the Niobe II MNS (Stereotaxis) was compared with manual ablation performed by the same operator in a different setting. The authors should be praised for the number of patients included (over 200) and their long follow-up (over three years on average). The study demonstrates that operator fluoroscopic exposure is clearly lower (mean of five minutes) using MNS, which is expected to lead to clear long-term benefits for the operator.

Remotely controlled ablation proved to be as safe as manual ablation. As there were concerns about lower contact force with the MNS, longer ablation times were observed in the MNS group. Whether or not this contributed to the lower relapse rate observed with MNS remains to be determined. However, the non-significant difference (due to the lack of statistical power) corresponds to an absolute risk difference of 3%, which is important, as it corresponds to 30 patients being referred for MNS ablation to avoid one relapse. This is a considerable number, as it will lead to further admissions to the emergency room, redo procedures, and prescriptions, which is important from the standpoint of health resource use. A cost-effectiveness study taking these points into account may be valuable.

This study's findings are thought-provoking. A future randomized study using contact force sensing (which is currently becoming the standard) that aims to answer these questions is warranted.

Conflicts of interest

The authors have no conflicts of interest to declare.

References

1. Haissaguerre M, Warin JF, Lemetayer P, et al. Closed-chest ablation of retrograde conduction in patients with atrioventricular nodal reentrant tachycardia. *N Engl J Med*. 1989;320:426-33.
2. Epstein LM, Scheinman MM, Langberg JJ, et al. Percutaneous catheter modification of the atrioventricular node. *Circulation*. 1989;80:757-68.
3. Lee MA, Morady F, Kadish A, et al. Catheter modification of the atrioventricular junction with radiofrequency energy for control of atrioventricular nodal reentry tachycardia. *Circulation*. 1991;83:827-35.
4. Page RL, Joglar JA, Caldwell MA, et al. 2015 ACC/AHA/HRS guideline for the management of adult patients with supraventricular tachycardia: executive summary: a report of the American College of Cardiology/American Heart Association Task Force on Clinical Practice Guidelines and the Heart Rhythm Society. *J Am Coll Cardiol*. 2016;67:1575-623.
5. Blomström-Lundqvist C, Scheinman MM, Aliot EM, et al., European Society of Cardiology Committee, NASPE-Heart Rhythm Society. ACC/AHA/ESC guidelines for the management of patients with supraventricular arrhythmias – executive summary. A report of the American College of Cardiology/American Heart Association Task Force on practice guidelines and the European Society of Cardiology Committee for practice guidelines (writing committee to develop guidelines for the management of patients with supraventricular arrhythmias) developed with NASPE-Heart Rhythm Society. *J Am Coll Cardiol*. 2003;42:1493-531.
6. Haïssaguerre M, Jais P, Shah DC, et al. Spontaneous initiation of atrial fibrillation by ectopic beats originating in the pulmonary veins. *N Engl J Med*. 1998;339:659-66.
7. Deisenhofer I, Zrenner B, Yin YH, et al. Cryoablation versus radiofrequency energy for the ablation of atrioventricular nodal reentrant tachycardia (the CYRANO Study): results from a large multicenter prospective randomized trial. *Circulation*. 2010;122:2239-45.
8. Swissa M, Birk E, Dagan T, et al. Radiofrequency catheter ablation of atrioventricular node reentrant tachycardia in children with limited fluoroscopy. *Int J Cardiol*. 2017;236:198-202.
9. Bailin SJ, Korthas MA, Weers NJ, et al. Direct visualization of the slow pathway using voltage gradient mapping: a novel approach for successful ablation of atrioventricular nodal reentry tachycardia. *Europace*. 2011;13:1188-94.
10. Steinberg BA, Piccini JP. High-density mapping of the tachycardia circuit in atrioventricular nodal reentrant tachycardia. *Heart Rhythm Case Rep*. 2016;2:451-3.
11. Choe WC, Sundaram S, Boorman C, et al. High-density mapping of the slow pathway in a patient with atrioventricular nodal reentry given intranasal Etripamil during the NODE-1 study. *Heart Rhythm Case Rep*. 2017;3:479-82.
12. Casella M, Dello Russo A, Pelargonio G, et al. Near zero fluoroscopic exposure during catheter ablation of supraventricular arrhythmias: the NO-PARTY multicentre randomized trial. *Europace*. 2016;18:1565-72.
13. Dalal AS, Nguyen HH, Bowman T, et al. Force-sensing catheters during pediatric radiofrequency ablation: the FEDERATION study. *J Am Heart Assoc*. 2017;6.
14. Parreira L, Marinheiro R, Carmo P, et al. Atrioventricular node reentrant tachycardia: remote magnetic navigation ablation versus manual ablation – impact on operator fluoroscopy time. *Rev Port Cardiol*. 2019;38:183-8.